



DEC 02 2002

**SPECIAL 510(k) SUMMARY**

**1.0 Submitter:**

Name: WRP Asia Pacific Sdn Bhd  
Address: Lot 1, Jalan 3, Kawasan Perusahaan Bandar Baru Salak Tinggi,  
43900 Sepang, Selangor Darul Ehsan, MALAYSIA  
Phone No.: +60 3 8706 1486  
Fax No.: +60 3 8706 1485

Date of Summary Prepared: **17 OCT 2002**

**2.0 Contact Person:**

Name: Mr. Terence Lim  
Phone No.: +60 3 8706 1486  
Fax No.: +60 3 8706 1485

**3.0 Modified Device Identification:**

Trade Name: 1) Dermagrip-P, and  
2) Multiple or Customers' Trade Name  
Device Name: Powder Free Pink Nitrile Examination Gloves, Non-Sterile  
Common Name: Examination Gloves  
Classification Name: Patient Examination Gloves (per 21 CFR 880.6250)

**4.0 Identification of the Legally Marketed Device:**

Class I patient examination gloves, 80LZA, powder free, that meets all the requirements of ASTM standard D 6319 – 00a<sup>E3</sup> and FDA 21 CFR 800.20.

**5.0 Description of Device Modification:**

The Powder Free Pink Nitrile Examination Gloves, Non Sterile is equivalent to the existing model, i.e. Powder Free Blue Nitrile Examination Gloves, Non Sterile which had submitted and cleared under 510(k) number K022904.

K023517



WRP Asia Pacific Sdn Bhd

147817V

The difference in this submission is:

- a) Change of colour additive used, i.e. Pink

The modification of colour does not affect the intended use of the device as well as it does not affect its safety and effectiveness. The indication for use and proposed labeling for the device are illustrated in subsequent sections.

The Powder Free Pink Nitrile Examination Gloves, Non-Sterile meets all the requirements of ASTM standard D 6319 – 00a<sup>E3</sup> and FDA 21 CFR 800.20.

#### 4.0 Intended Use of the Device:

The Powder Free Pink Nitrile Examination Gloves, Non-Sterile is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

#### 5.0 Summary of Technological Characteristics for the Modified Device:

The Powder Free Pink Nitrile Examination Gloves, Non-Sterile are summarized with the following technological characteristics compared to ASTM or equivalent standards.

CHARACTERISTICS	STANDARDS	DEVICE PERFORMANCE
Dimensions	ASTM D 6319 – 00a <sup>E3</sup>	Meets
Physical Properties	ASTM D 6319 – 00a <sup>E3</sup>	Meets
Freedom from pinholes	ASTM D 6319 – 00a <sup>E3</sup> FDA 21 CFR 800.20	Meets
Powder Residual	ASTM D 6124 – 01	Meets < 2 mg/glove

#### 6.0 Conclusion:

The Powder Free Pink Nitrile Examination Gloves, Non-Sterile will perform according to the glove performance standards referenced in section 7 above and meet ASTM standards, and FDA requirements for waterleak test on pinhole AQL. Consequently, this device is substantially equivalent to currently marketed devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

DEC 02 2002

Mr. Terence Lim  
Associate Manager, RA/QA  
WRP Asia Pacific Sdn. Bhd.  
Lot 1, Jalan 3, Kawasan Perusahaan  
Bandar Baru Salak Tinggi,  
43900 Sepang,  
Selangor Darul Ehsan,  
MALAYSIA

Re: K023517

Trade/Device Name: DermaGrip-P Powder Free Pink Nitrile Examination  
Gloves, Non-Sterile  
Regulation Number: 880.6250  
Regulation Name: Patient Examination Gloves  
Regulatory Class: I  
Product Code: LZA  
Dated: November 7, 2002  
Received: November 12, 2002

Dear Mr. Lim:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4618. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

A handwritten signature in dark ink, appearing to read "Susan Runner" or similar, is written over the typed name "Timothy A. Ulatowski".

Timothy A. Ulatowski  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure



WRP Asia Pacific Sdn Bhd

147817V

### INDICATIONS FOR USE

Applicant: WRP Asia Pacific Sdn Bhd

510(k) Number (if known): Ø 23517

Device Name: DERMAGRIP-P POWDER FREE PINK NITRILE  
EXAMINATION GLOVES, NON-STERILE

#### Indications For Use:

The Powder Free Pink Nitrile Examination Glove, Non-Sterile is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

\_\_\_\_\_  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use \_\_\_\_\_ OR Over-The-Counter ☒  
(Per 21 CFR 801.109)

Susan Runov  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital,  
Infection Control, Dental Devices

510(k) Number K023517